

# BLANKET PURCHASE ORDER

## STATE OF MARYLAND

\*\*\*\*\* STATE OF MARYLAND \*\*\*\*\*

BPO NO: 001B3400381

PRINT DATE: 10/31/15

PAGE: 01

<b>SHIP TO:</b> AS SPECIFIED ON INDIVIDUAL ORDERS		
<b>VENDOR ID:</b> TRINITY BIOTECH DISTRIBUTION 2823 GIRTS RD PO BOX 1059 JAMESTOWN, NY 14702-1059 (716 )483-3851	<b>REFER QUESTIONS TO:</b> IRIS LESTER (410 )767-4612 IRIS.LESTER@MARYLAND.GOV	
<b>ITB:</b> M00R3401506	<b>EXPR DATE:</b> 01/14/19 <b>POST DATE:</b> 02/15/13	<b>DISCOUNT TERMS:</b> . NET 30 DAY <b>CONTRACT AMOUNT:</b> .00

### TERMS:

ARTICLES HEREIN ARE EXEMPT FROM MARYLAND SALES AND USE TAXES BY EXEMPTION CERTIFICATE NUMBER 3000256-3 AND FROM FEDERAL EXCISE TAXES BY EXEMPTION NUMBER 52-73-0358K. IT IS THE VENDOR'S RESPONSIBILITY TO ADVISE COMMON CARRIERS THAT AGENCIES OF THE STATE OF MARYLAND ARE EXEMPT FROM TRANSPORTATION TAX.

<u>LINE #</u>	<u>STATE ITEM ID</u>	<u>U/M</u>	<u>UNIT COST</u>
0001	41578-000021	EA	6.2700

LABORATORY MISC. SUPPLIES

TWO ULTRA 2 HPLC GENESYS NBS ANALYZERS, CAT.# 03-01-0045  
INCLUDES: INSTRUMENT, SERVICE, ON SITE INSTRUMENT OPERATION AND  
PROFILE TRAINING, REAGENTS AND CONSUMABLES.  
WILL BE INVOICED MONTHLY FOR THE NUMBER OF REPORTABLE TESTS GENERATED  
EACH MONTH. PRICING BASED ON 6250 REPORTABLE PATIENT TESTS/YR.  
COST PER REPORTABLE: \$ 6.27 FOR FIVE YEARS.  
5 YEAR BPO TO BEGIN 01/01/13 THRU 12/31/18  
TECHNICAL SPECIFICATIONS FOR NBS HEMOGLOBINOPATHIES  
- THE COLUMN AND REAGENTS MUST BE THE SAME FOR NEWBORN AND ADULT  
SAMPLES  
- INSTRUMENT MUST BE CAPABLE OF AUTOMATED SAMPLE PREPARATION FOR PAPER  
SPOTS AND PRIMARY TUBES TO INCLUDE AUTO-DILUTION OF MICROTITRE  
PLATES  
- COLUMN AND REAGENT COMBINATIONS MUST BE NON-LOT SPECIFIC AND  
INTERCHANGEABLE  
- INSTRUMENT MUST USE RELATIVE RETENTION TIMES FOR THE IDENTIFICATION  
OF HB VARIANTS  
- COLUMNS MUST BE COMPOSED OF POLYCAT A WEAK CATION EXCHANGE MATERIAL  
- MOBILE PHASES MUST UTILIZE BIS-TRIS BUFFERS AT PH 6.8 AND 6.6

\*\*\* CONTINUED, NEXT PAGE \*\*\*

# BLANKET PURCHASE ORDER

## STATE OF MARYLAND

\*\*\*\*\* STATE OF MARYLAND \*\*\*\*\*

**BPO NO:** 001B3400381

**PRINT DATE:** 10/31/15

**PAGE:** 02

<u>LINE #</u>	<u>STATE ITEM ID</u>	<u>U/M</u>	<u>UNIT COST</u>	_____	_____
---------------	----------------------	------------	------------------	-------	-------

- INSTRUMENT MUST USE A DETECTION OF 413NM
- HB F, A, S, C, D, O, E AND BART'S MUST BE DETECTED, LABELED AND QUANTITATED
- ON BOARD SAMPLE CAPACITY OF 850 PER BATCH
- MINIMUM READABLE DIVISION; AREA + 0.05%, RETENTION TIME = 0.001 MINUTES
- INSTRUMENT COMMUNICATION MUST BE WINDOWS XP BASED
- THROUGHPUT RATE: 2 MINUTES FOR QUICKSCAN OF NEONATAL SAMPLES; 8 MINUTES FOR HIGH RESOLUTION CONFIRMATION OF NEONATAL SAMPLES OR QUANTITATION OF A2, F AND HB VARIANTS IN ADULT SAMPLES

\_\_\_\_\_ END OF ITEM LIST \_\_\_\_\_

\*\*\*\*\* LAST PAGE \*\*\*\*\*

**AUTHORIZED BY:** \_\_\_\_\_ **DATE:** \_\_\_\_\_

BUYER AUTHORIZED DESIGNEE